



AIR/GAS QUALITY REPORT & CERTIFICATE


The air/gas analyzed on 04/10/15 satisfies the guidelines for
CGA E (G-7.1, 2011) / CGA D (G-7.1, 2011)
ALKIN KOMPRESSOR SAN. VE TIC. LTD

SN: W4 & 948204

Client ID: NA Agent ID: ALK135470 ALKIN KOMPRESSOR SAN. VE TIC. LTD

Kit #: 309350	Frequency: Single		
Analytes	04/10/15	Range	
	282497.SPA		
Oxygen (%)	21.4	20-22	
Carbon Dioxide (ppm)	<50	1000	
Carbon Monoxide (ppm)	<1	10	
T.H.C. (ppm)	<1	25	
Water Vapor (ppm)	18.81	67	
Dew Point (F)	-68	-50	
Oil/Part. (mg/m3)	<1	5	
Unknowns	ND	NA	
Odor	ND	None	

Results: Sample Meets Standard



IMPORTANT: This air/gas or substance was tested ONLY against a specified standard and may contain undetected items which are beyond the purpose or scope of this analysis. More extensive testing can be conducted upon request. This does NOT guarantee the condition nor safe application of the analyzed air/gas or substance. Results reported relate only to the items tested. This report shall not be reproduced except in full, without the consent of Lawrence Factor, Inc. Methods used: FTIR, GC, electrochemical, microgravimetry, and sniffing. Estimated Measurements of Uncertainty values for the reported analytes are available upon request. This report / certificate is not issued for an unbounded duration. Kit#: 309350 - Frequency: Single - Date Received: 04/08/15 - Issue ID: LF282497 - Page 1 of 1



[Signature]
Endorsement / President

[Signature]
Endorsement / Lab Director